

NABL ACCREDITED LABORATORY FOR CALIBRATION OF MEDICAL EQUIPMENT

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP, SPO ₂ , INVASIVE PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP LEAK TEST, NIBP DYNAMIC, SPO ₂	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	PATIENT MONITOR	ECG, NIBP, SPO ₂ , TEMPERATURE, RESPIRATION	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	PATIENT MONITOR	HEART RATE, RESPIRATION RATE, NIBP LEAKAGE TEST, NIBP TEST, NIBP PRESSURE RELIEF, INVASIVE BLOOD PRESSURE, SPO ₂ , HEART RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5.	PATIENT MONITOR	IABP TEMPERATURE RESPIRATION RATE HEART RATE NIBP DYNAMIC NIBP LEAK TEST	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	VENTILATOR	TIDAL VOLUME, RESPIRATORY RATE, MEAN AIRWAY PRESSURE, POSITIVE END EXPIRATORY PRESSURE (PEEP), OXYGEN CONCENTRATION	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	VENTILATOR	TIDAL VOLUME, FIO ₂ , BREATH RATE, I:E RATIO, PEEP	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	VENTILATOR	VOLUME ACCURACY, RESPIRATION RATE, INSPIRATORY EXPIRATORY TIME, PRESSURE ACCURACY (PIP), PEEP, OXYGEN PERCENTAGE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE ROAD AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	INFUSION PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	INFUSION PUMP	FLOW RATE ACCURACY, VOLUME ACCURACY, OCCLUSION DETECTION PRESSURE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	INFUSION PUMP	FLOW, VOLUME, OCCLUSION PRESSURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION DETECTION PRESSURE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	SYRINGE PUMP	FLOW RATE, VOLUME, OCCLUSION PRESSURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, DISCHARGE TIME, SYNCHRONIZER OPERATION	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, ENERGY AFTER 10 CONSECUTIVE DISCHARGES, TIMELY CHARGING	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	DEFIBRILLATOR	ECG, ENERGY, SYNC TEST	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM,		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	DEFIBRILLATOR	HEART RATE, OUTPUT ENERGY, OUTPUT ENERGY AT MAXIMUM SETTING FOR 10CHg CYCLE (AT BATTERY POWER), CHARGE TIME AFTER 10 DISCHARGE CYCLES, ENERGY AFTER 60 SEC OF FULL CHARGE AT MAX ENERGY, SYNCHRONIZER OPERATION, NIBP TEST, O ₂ SATURATION SPO ₂ (HEART RATE), PACER AMPLITUDE, PACER RATE ACCURACY	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	DEFIBRILLATOR	OUTPUT ENERGY, DISCHARGE TIME (POSITIVE / NEGATIVE PULSE WIDTH), HEART RATE, SYNCHRONIZATION TEST-CHARGE TIME	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	SUCTION PUMP/ ASPIRATOR	VACUUM (PRESSURE)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	SUCTION PUMP	PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO- ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	SUCTION PUMP	VACUUM	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080- 23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BP APPARATUS	LEAK TEST NIBP (STATIC PRESSURE)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	BP APPARATUS	NIBP LEAK TEST	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	BP APPARATUS	PRESSURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	BP APPARATUS	SPHYGMOMANOMETER: LEAK TEST GAUGE ZERO PRESSURE ACCURACY DIGITAL BP APPARATUS: LEAK TEST PRESSURE ACCURACY HEART RATE ACCURACY	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	ECG MACHINE	HEART RATE AMPLITUDE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	ECG MACHINE	HEART RATE AMPLITUDE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	ECG MACHINE	HEART RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO- ORPORATION, TRIVANDRUM TALUK,		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	ECG MACHINE	HEART RATE, AMPLITUDE,	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080- 23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
5	ECG MACHINE	HEART RATE, AMPLITUDE PATIENT LEAD LEAKAGE CURRENT, CHASIS LEAKAGE, GROUND WIRE RESISTANCE, INSULATION RESISTANCE, PATIENT LEAKAGE,	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	NEBULIZER	FLOW RATE	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	NEBULIZER	FLOW RATE	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	NEBULIZER	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT		9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
4	NEBULIZER	FLOW RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	RADIANT WARMER AND INFANT INCUBATOR	TEMPERATURE OF WARMER, HUMIDITY IN WARMER, AIR FLOW, SOUND LEVEL	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	RADIANT WARMER	TEMPERATURE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	RADIANT WARMER	TEMPERATURE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
4	RADIANT WARMER , PATIENT WARMER	TEMPERATURE	PRISM CALIBRATION CENTRE	GF-101,F/101,101 A,B, TF-85 TO 101 RUDRAKSH COMPLEX-II, JASHODA NAGAR CROSS ROADS, PHASE-III GIDC, VATVA, AHMEDABAD,	GUJARAT	9428602063	info@prismcalibration.com	KRISHNA KINARIWALA

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	PULSE OXYMETER	HEART RATE SpO2	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	PULSE OXYMETER	HEART RATE SpO2	CYRIX HEALTH CARE PRIVATE LIMITED, CALIBRATION LABORATORY DIVISION,	NO. 30/641B, PETTA JUNCTION, POONITHURA, KOCHI	KERALA	9744400633	calibrationcyrix@gmail.com	MS. RAKHI S. NAIR
3	PULSE OXYMETER	HEART RATE SpO2	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KARNATAKA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K

S. NO.	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BOYLES APPARATUS	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
2	BOYLES APPARATUS	FLOW RATE	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD, MALLESWARAM, BANGALORE	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR

S. NO	PRODUCT	PARAMETERS IN NABL ACCREDITED SCOPE	LABORATORY	ADDRESS	STATE	PHONE NO.	EMAIL	CONTACT PERSON
1	BLOOD GAS ANALYZER, SEMI AUTO ANALYSER, HEMATOLOGY ANALYZER, O.T. TABLE,	GROUND WIRE RESISTANCE, CHASIS LEAKAGE CURRENT, PATIENT LEAKAGE CURRENT, PATIENT LEAD LEAKAGE CURRENT, ISOLATION TEST (MAINS ON APPLIED PARTS) INSULATION TEST (OPTIONAL 500V)	APEX QUALITY CERTIFICATION SERVICES PVT. LTD.	161, KASTURBA NAGAR, NIRMAN NAGAR, JAIPUR	RAJASTHAN	9414079483	ratiram_yadav@yahoo.co.in	MR. R R YADAV
2	FLOW METER	FLOW RATE	INSTITUTE FOR TESTING & ANALYSIS OF BIOMEDICAL INSTRUMENTS	PNRA-F3, SK LANE, KOWDIAR P.O., TRIVANDRUM CO-ORPORATION, TRIVANDRUM TALUK, TRIVANDRUM DISTRICT	KERALA	9495746216	info@intab.org	MR. VIKRAMAN PILLAI K
3	BiPAP	IPAP, EPAP, RR, Ti	TRANSCAAL ENGINEERS INDIA PVT. LTD.	#96/1-1, 1ST FLOOR, EAST PARK ROAD, 9TH CROSS, BETWEEN SAMPIGE AND MARGOSA ROAD,	KARNATAKA	080-23362266	mail@tepl.info	MR. G.K. SURESH KUMAR
4	AUTOCLAVE	TEMPERATURE PRESSURE GAUGE		MALLESWARAM, BANGALORE				